FS-300 Thermofinder

NON-CONTACT Infrared Thermometer





















[Normal]



Temperature readings may vary according to location of measurement from forehead to temple Therefore we recommend to measure the temperature more than 2 points on the forehead. Then please recognize the highest temperature as the right body temperature.

[Sweating on the forehead]



We recommend changing of measurement point to behind the ear lobe from forehead and temple. Because this area is less affected by sweating than the forehead.



SPECIFICATIONS

OF LOW TONG	
Measurement Method	Infrared measurement, Non-Contact measurement
Measurement distance	2 to 3cm from the measuring point
Measurement point	The measuring point is the temple area
Measurement range	Body temperature: 34°C to 42.5°C (93.2°F to 108.5°F)
	Accuracy ±0,2°C (36°C to 39°C)
	Other Accuracy ±0.3℃
	Object temperature: 15°C to 60°C (59°F to 140°F)
	Accuracy ±2°C
Response time	Less than 2 seconds continuous measurement
Resolution	0.1°C
Memory	Automatically body temperature data (maximum 32) stored
Back-Light	If measurement complete, lighted 3 seconds
Automatic Power off	Auto Power-Off (after 1 minute)
Accessory	FS-300 1set, Holder 1ea, Manual 1volume, Battery(AAA type) 2ea
Power supply	DC3V(AAA 1.5V Battery 2ea)

