

Lite Tech Inc.

975 Madison Avenue
Norristown
PA 19403,
USA..

It is certified that the manufacturer's technical file and the PPE product detailed on page 2 have been assessed and found to be in accordance with

EC Directive 89/686/EEC

EC Type-examination

This certificate is valid from 17 April 2014
Issue 1. Certified since 28 September 2007
This supersedes certificates GB07/73536, GB07/73537, GB07/73538,
GB07/73539, GB07/73540 & GB07/73541.


Authorised by

SGS United Kingdom Limited, Notified Body 0120
Unit 202B Worle Parkway, Weston-super-Mare, BS22 6WA, United Kingdom
t +44 (0)1934 522917f +44 (0)1934 522137
e-mail sgsprodcert@sgs.com www.sgs.com

SGS EC 03 0311

This certificate remains the property of SGS United Kingdom Ltd to whom it must be returned on request.

Page 1 of 3



Lite Tech Inc.

EC Directive 89/686/EEC

EC Type-examination

Issue 1.

PPE Product

Lite Tech "Xenolite NLE", X-ray protective clothing providing protection to parts of the body with various nominal lead equivalences of at 100kV. This clothing is available in a range of sizes and styles (product codes as follows):

With nominal lead equivalences of 0.25mm, 0.35mm or 0.50 mm in each panel at 100kV;

Frontal protection only: 810B, 810S, 810A, 810Q, 810K, 810-NV, 810E, 810R, 840H & 810PHP.

Accessories: 850H and 840T & 840T-BC (840T & 840T-BC in 0.50mm only and available in only one size).

With nominal lead equivalences of 0.25mm at 100kV on the front & 0.25mm at 100kV on the back;

Front & back protection: 810-L, 8201, 8250, 8200TS, 820V, 820S, 830V, 830S, 835V & 835S

With nominal lead equivalences of 0.35mm or 0.50mm at 100kV on the front & 0.25mm at 100kV on the back;

Front & back protection: 810-L, 8201, 8240, 8240-NV, 8250, 8200TS, 8320, 820V, 820S, 825V, 825S, 830V, 830S, 835V & 835S.

With nominal lead equivalences of 0.35mm at 100kV on the front & 0.35mm at 100kV on the back;

Front & back protection: 8200TS, 8201, 8240, 8240-NV, 8250, 8320, 820V, 820S, 825V, 825S, 830V, 830S, 835V & 835S.

With nominal lead equivalences of 0.50mm at 100kV on the front & 0.35mm at 100kV on the back;

Front & back protection: 8201, 8250, 8200TS, 820V, 820S, 830V, 830S, 835V & 835S.

All of the garments must be worn following the manufacturer's instructions for use to obtain the claimed level of protection.

Continued.....

Lite Tech Inc. **EC Directive 89/686/EEC** **EC Type-examination**

Issue 1.

It is certified that the manufacturer's technical file and the above mentioned PPE have been assessed and found to be in accordance with the requirements of Council Directive 89/686/EEC. When examined the PPE satisfied the requirements in the manufacturer's technical specification¹ (based on EN 61331-1:2002 test requirements), which had been previously assessed and deemed to meet the basic health and safety requirements (Annex II) of the Directive.

The manufacturer's information notice ((LITE TECH "XENOLITE" B, NL, TB, S & "EVAL" X-RAY PROTECTION GARMENTS & ACCESSORIES INFORMATION NOTICE TO USERS ISSUE # 6 02/2014)), in English, has been inspected and found to have addressed all of the relevant requirements of the standard. However, the detailed content of the information notice is the responsibility of the manufacturer, as are translations into other languages.

This certificate is issued on the strict condition that appropriate checks on manufactured PPE, as detailed in Article 11 of the Directive are implemented and maintained while the model is in production.

Certification is based on technical file specifications: Lite Tech Tech File Version 6
dated: 17 Feb 2014

SGS Reference Number GB/PP 211441

¹ The validity of this certificate may have to be reviewed in the light of new or amended harmonized standards.

This certificate remains the property of SGS United Kingdom Ltd to whom it must be returned on request.